

CERTIFICATE OF CALIBRATION

Certificate #: 90048

Keep for your records.

<u>Customer:</u>	Sample 123 Any Street Any Town, CA 12345 United States	Laboratory Location:	Max Machinery, Inc 33A Healdsburg Ave Healdsburg, CA 95448 707-433-2662
<u>Type of Device:</u> <u>Manufacturer:</u> <u>Model Number:</u> Serial Number:	Flow Meter Max Machinery, Inc. 234-100-000, D12345,	<u>Calibration Fluid:</u> Fluid Viscosity: Fluid Specific Gravity: Fluid Temp:	Water 1 cps 1.00 g/mL 21°C +/- 1°C
Date of Calibration: Sales Order: Procedure Used: Performed By:	7/1/2015 SAMPLE LA-P-XXX JDO	<u>Output Units:</u> <u>Flow Units:</u> Ave K-Factor:	Pulses/mL mL/min 45.887
Calibration Notes:	This document reflects the new basic calibration.		

This document reflects the new basic calibration. The new condition was found to be in tolerance.

Calibration Data

Flow Rate	Output		Error	Flow	Flow Rate Output		
mL/min	Pulses/mL	Frequency Hz	% reading	n	nL/min Pulses/mL	Frequency Hz	% reading
4000.0	45.574	3038.267	-0.68%				
3000.0	45.633	2281.650	-0.55%		-11		
1000.0	45.979	766.317	0.20%				
500.0	46.165	384.708	0.61%	DATOP	VI		
300.0	46.201	231.005	0.68%	MICH			
100.0	45.993	76.655	0.23%				
20.0	44.985	14.995	-1.97%				
10.0	44.063	7.344	-3.97%		A		
			× 001:200				
				°			

Average K-Factor is calculated as a weighted average over the nominal range of the meter.

Equipment Used in the Calibration:

Calibration ID: 40801

Description: Reference Standard 1 Serial Number: 12345

Cal Due Date: 3/13/2016

Certificate Number: sampleCert

QC Approval:

ne Doe 7/1/2015

7/1/2015

John Doe Lab Technician

Calibration Technician

Jane Doe **Quality Manager**

This Certificate shall not be reproduced, except in full, without written approval by Max Machinery, Inc.