

CERTIFICATE OF CALIBRATION

Certificate #: 90235

Keep for your records.

| Customer: | Sample Customer | Laboratory Location: | Max Machinery, Inc |
|----------------------|---------------------|-------------------------|----------------------|
| | 123 Any Street | | 33A Healdsburg Ave |
| | Any Town, CA 12345 | | Healdsburg, CA 95448 |
| | United States | | 707-433-2662 |
| | | Calibration Fluid: | Kerosene |
| Type of Device: | Flow Meter | Fluid Viscosity: | 3 cps |
| Manufacturer: | Max Machinery, Inc. | Fluid Specific Gravity: | 0.82 g/mL |
| Model Number: | P214MS11NA/P11N/0 | Fluid Temp: | 21°C +/- 1°C |
| Serial Number: | D12345 | | |
| Date of Calibration: | 7/1/2015 | Output Units: | Pulses/mL |
| Sales Order: | SAMPLE | Flow Units: | mL/min |
| Procedure Used: | LA-P-113 | K-Factor: | 90.000 |
| Performed By: | JDO | | |

<u>Calibration Notes:</u> This document reflects the new linear calibration. The new condition was found to be in tolerance.

Calibration Data

| Flow Rate | Out | put | Error | Flow Rate Output | Error |
|-----------|-----------|--------------|--|-------------------------------|-----------|
| mL/min | Pulses/mL | Frequency Hz | % reading | mL/min Pulses/mL Frequency Hz | % reading |
| 10042.8 | 89.987 | 15062.020 | -0.01% | | |
| 7002.8 | 90.010 | 10505.370 | 0.01% | | |
| 2999.3 | 90.018 | 4499.850 | 0.02% | | |
| 997.5 | 90.026 | 1496.682 | 0.03% | PATOPY | |
| 299.9 | 90.039 | 450.045 | 0.04% | | |
| 100.0 | 90.042 | 150.070 | 0.05% | | |
| 30.0 | 90.012 | 45.006 | 0.01% | | |
| 10.0 | 90.052 | 15.009 | 0.06% | | |
| | | | ~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~~ | | |
| | | | | 08 | |

Equipment Used in the Calibration:

| Calibration ID: | Description: | Serial Number: | Cal Due Date: | Certificate Number: |
|-----------------|--------------------------|----------------|---------------|---------------------|
| 41801 | Built in reference meter | 418 | 3/13/2016 | 41801031315 |
| 41802 | Built in reference meter | 418 | 3/13/2016 | 41802031315 |
| 41803 | Temperature controller | 41803 | 11/12/2015 | 41803111214 |
| 41805 | Counter/timer | 166FC5C | 11/15/2015 | 41805111514 |
| 41807 | Multifunction DAQ | 1676A16 | 11/26/2015 | 41807112614 |

QC Approval:

ne Doe 7/1/2015

Calibration Technician

7/1/2015 John Doe

Lab Technician

Jane Doe

Quality Manager

This calibration was conducted using standards traceable to NIST.

Measurement uncertainty of the #418 test stand is +/- 0.126% of reading with a 95% confidence (k=2 coverage factor). Calculations are available upon request.

This Certificate shall not be reproduced, except in full, without written approval by Max Machinery, Inc.

Max Machinery, Inc. maxmachinery.com 33A Healdsburg Avenue Healdsburg, CA 95448

T +1 707.433.2662 F +1 707.433.1818 Print Date: 23 Nov 2016 1:15:00 PM Page 1 of 1